

HPLC-Chip/MS

**A new microfluidic platform for
easy to use, robust and reliable
nanospray LC-MS**



Agilent Technologies, LSCA

Tom A. van de Goor, Ph.D.

R&D and Marketing GmbH&Co KG

Email: Tom_Vandegoor@Agilent.com

Prague, September 26, 2007



Agilent in Life Sciences and Chemical Analysis

\$20 billion total addressable market... ample opportunity for growth

Life Sciences


Pharma, Biotech,
CRO, CMO



Academia &
Government



Clinical
Diagnostics DX



Addressable Market: \$14B
Market Growth Rate: 7-9%
Agilent Projected Grpwth: 12-15%



Chemical Analysis

Petrochemical
QA/QC



Environmental
Testing



Food
Testing





Forensics
Testing



Addressable Market: \$6B
Market Growth Rate: 5-7%
Agilent Projected Growth: 7-9%



Life Sciences Tools



Consumables



Spectroscopy Tools



Separations Tools




Services & Informatics



Agilent Technologies

HPLC-Chip/MS

New Agilent High Performance 1200 Series LC: Best in Class “Front End” Systems for MS



Preparative LC

Standard LC

**Rapid Resolution
LC**

**Capillary &
Nano LC**

**High Sensitivity
HPLC-Chip/MS**

*The industry's most comprehensive LC portfolio to configure an
integrated LC/MS system*



6000 Series LC/MS Portfolio

New Platforms



6410 Triple Quad LC/MS

- Femtogram sensitivity
- Excellent reproducibility,
- Cost of ownership advantages

Food, Forensic, Environmental, Chemical,
Pharmaceutical, Proteomics, Metabolomics



6510 QTOF LC/MS

- 10X sensitivity advantage
- 2X typical mass accuracy
- Attomole MS and MS/MS
- Wide in-scan dynamic range
- Fast scan speeds (20/s)

Proteomics, Metabolomics,
Pharmaceutical



6000 Series LC/MS Portfolio

Enhanced Platforms



6300 Series Ion Trap LC/MS

Speed, resolution, range, sensitivity

- MS and MSⁿ Workhorse
- 6340: Electron Transfer Dissociation provides soft ion fragmentation
- Use on chromatographic time scale

Proteomics, Pharmaceutical, Food Safety, Environmental, Clinical



6210 TOF LC/MS

“FTMS Like” Performance

- 2 ppm Mass Accuracy
- High mass resolution
- Wide dynamic range (3-4 orders)
- Fast scan speeds for high throughput screening

Food, Forensics, Environmental, Chemical, Pharma, Metabolomics



Another New LC/MS in the 6000 Series

NEW
AT
ASMS

6100 Series Single Quads

Higher Performance

6140 handles 1200 Rapid Resolution LC data

Greater Choice and Full Investment Protection

4 Models, Each One Upgradable to the Next

Better Price/Performance on All Models

40% smaller footprint



- 6140:
1-picogram sensitivity, fast-scan
to 10,000 amu/second
- 6130, 6120, 6110:
High performance to entry level
models,
1 to 10 picogram sensitivity,
Mass range to 3,000 amu

**Food, Forensics, Environmental Chemical,
Pharmaceutical**



Agilent Technologies

HPLC-Chip/MS

First Lab-on-a-Chip Product




Caliper Lab-Chip Kits




Agilent 2100 Bioanalyzer

Proteins in human serum



A Big Biological Challenge

In human serum the concentrations of proteins range over twelve orders of magnitude



Current technologies permit one to only look at the top layers: **High abundance proteins**

A major goal is to look at clinically useful proteins (biomarkers) which are typically **Moderate to Low abundance proteins**.

Outline

Liquid Chromatography and hyphenated techniques

- Liquid chromatography, detection
- Hyphenated techniques
- Miniaturizing HPLC-MS

The Agilent HPLC-Chip/MS system

- HPLC-Chip manufacturing and chip layouts
- HPLC-Chip system hardware components
- Key performance parameters

Application examples


Outlook

- Future applications and functionalities



Liquid chromatography

For separation of compounds in solution



How it used to be done



Agilent 1200 Series HPLC

Slide courtesy of Jim Hollenhorst, Agilent Laboratories



High Performance Liquid Chromatography

HPLC is an analytical technique for the **separation** of compound mixtures as a function of time based on *multiplicative distribution* of the sample component between a moving *mobile phase* and a stagnant *stationary phase*

Well established in:

- Environmental analysis
- Pharmaceutical analysis
- Bioanalysis (Genomics, Proteomics and Metabolomics)
- Food analysis
- Quality control
- Process control



HPLC is universal and of superior significance in routine analysis as well as research



Detectors for HPLC

A HPLC detector locates, in the dimension of time, the sample components that have been subjected to a chromatographic process.

Optimal characteristics:

- **High sensitivity**
 - Defined response for quantitative analysis
 - Broad applicability
 - Unaffected response to chromatographic parameters
 - No contribution to extra-column band broadening
-
- Provides reliable qualitative information upon the separated analytes


eg. UV detectors
Fluorescence detectors
RI detectors
Conductivity detectors
...

Mass Spectrometry detectors



Pharmaceutical example

Identify and quantify impurities in drug manufacturing



Slide courtesy of Jim Hollenhorst, Agilent Laboratories




Agilent Technologies

HPLC-Chip/MS

Hyphenated techniques: HPLC - MS

Hyphenation techniques comprises all methods for ***coupling*** HPLC with ***detection units***, which provide any ***qualitative information*** upon the eluting sample components, eg:


- **HPLC-MS**
- HPLC-NMR
- HPLC-ICP-MS
- ...



The HPLC system and predominately the separation column properties have to be adapted to the respective detection device.

Why miniaturize HPLC/MS?

- Radically reduces sample volume
- Decrease solvent and stationary phase consumption
- Increase mass sensitivity with decreasing column I.D.
- Increase column efficiency with decreasing column I.D.
- Faster reaction time in smaller volumes (surface to volume ratio)



AND

there is a clear trend in the (Bio)Pharmaceutical industry and Life Science community to put advanced equipment like LC-MS, in the hands of research workers who are not liquid chromatography or MS specialists

➔ Strong need for easy-to-use & robust analytical tools




Why use microfluidics?



	Nano LC	Capillary LC	Micro LC	HPLC	Prep LC		
Flow Rate (μL/min)	0.1	1	10	100	1,000	10,000	100,000
Column i.d. (mm)	0.05	0.2	1	2	4	50	
Sensitivity gain	6400	400	16	4	1		



Achieve Sensitivity by Reduction of ID (2r)




$$\text{Intensity} = f(c) = \frac{m}{V_{sp}}$$



Nanospray LC/MS

Nanospray LC/MS combines both **high resolution chromatography** and **high sensitivity MS detection** and is used for ultra-trace analysis applications but...

Nanospray LC/MS is **challenging to implement and maintain**: multiple small capillary tubing connections, frequent clogging and leaks at the columns and nanospray tip and chromatographic degradation caused by tubing dead volume compromise reliability, ease of use, robustness and chromatographic performance.



Why use Microfluidics for nanospray LC/MS?


Microfluidic chip devices are designed to be used at nano and pico flow rates and can integrate most functional components of a conventional nanocolumn LC/MS directly on to a chip.

A microfluidic **HPLC-Chip** can integrate enrichment and analytical nanocolumns, nanospray emitter, fittings and connection capillaries directly on a reusable biocompatible polymer chip.

An automated HPLC-Chip interface ensures solvent and sample delivery to the chip, high pressure flow switching and automated chip loading and positioning in front of the MS inlet.



HPLC-Chip platform for nanospray LC/MS



HPLC-Chip/MS interface



1200 NanoLC System


6000 Series Mass Spectrometer
(Ion trap, SQ, QQQ, TOF, Q-TOF)




How Do You Make an HPLC-Chip?

Biocompatible Polyimide Films


- Films 25 - 150 μm
- Special brand with high solvent resistance
- Multiple layers to make a HPLC-Chip





Rapid and High-Precision Fabrication of Polymer Microfluidics using Direct Write Laser Ablation






Laser Ablated Microfluidic Elements




Channels





ES Tips



UV Window





Holes




Sippers



Masks



Laser Ablated Polyimide Electrospray Tips




20 µm




20 µm

Visualization of Electrospray





HPLC-Chip design features



HPLC-Chip/MS system hardware components

- Putting it all together



Putting it all together: how does it work?

- Easy to use HPLC-Chip system


1 Set up the nanoLC flow and gradient




2 Insert HPLC-Chip




3 Click on the Operate command in Chemstation



4 HPLC-Chip automatically moves into spray position




Chromatographic Performance




The integrated HPLC-Chip eliminates dispersion and band broadening and increases chromatographic performance and MS sensitivity.


Chromatographic Performance: HPLC-Chip Vs. Nanocolumn



Nanospray performance



Use of the orthogonal dual electrode nanospray ion source shows constant signal-to-noise for all types of Taylor cone plumes when combined with the polymer chip



- HPLC-Chip delivers robust spray performance for all solvent conditions.
- Does not require voltage change or optimization during a gradient.
- Polyimide nanospray emitter lifetime measured in weeks!


Data courtesy of Paul Goodley, Agilent Technologies (USA)



Agilent Technologies

HPLC-Chip/MS

Retention Time reproducibility



**Extracted ion chromatograms for 17 peaks
from a single analysis of BSA tryptic
digest (50 fmol on-column).
Rapid gradient method**

	Average RT	SD	%RSD
EIC 487.8 ±All MS	3.618	0.014	0.40
EIC 752	3.788	0.011	0.29
EIC 740.6 ±All MS	5.018	0.010	0.20
EIC 874.4	3.968	0.012	0.31
EIC 653.6	4.289	0.012	0.28
EIC 511.7	3.681	0.012	0.31
EIC 722.7	3.547	0.012	0.35
EIC 778	4.143	0.010	0.23
EIC 526.3	4.399	0.015	0.34
EIC 547.5	4.472	0.011	0.25
EIC 746.7	5.196	0.011	0.20
EIC 519.1	4.142	0.011	0.26
EIC 508.2	4.972	0.011	0.23
EIC 582.4	4.679	0.011	0.23
EIC 461.9	3.905	0.012	0.30
EIC 474	4.759	0.011	0.22
EIC 628	4.584	0.010	0.22

**RT reproducibility evaluated using 69 repeat injections
Worst case variation was 0.4% RSD, typical 0.25% RSD**



HPLC-Chip/MS



Application examples




Sensitivity of the HPLC-Chip/MS System: Results for 100 amol HSA Digest

Group (#)	Spectra (#)	Distinct Peptides (#)	Distinct Summed MS/MS Search Score	% AA Coverage	Mean Peptide Spectral Intensity	Database Accession #	Protein Name
1	6	6	83.98	12	1.40e+006	IP100022434	Serum albumin


#	Sequence	m/z Measured (Da)	MH ⁺ Matched (Da)	A100-1 Intensity
1	(K)FQNALLVR(Y)	480.97	960.5631	1.80e+006
2	(K)KVPQVSTPTLVEVSR(N)	547.47	1639.9383	1.50e+006
3	(K)LVNEVTEFAK(T)	575.46	1149.6156	1.93e+006
4	(K)QNCELFEQLGEYK(F)	829.49	1657.7532	1.04e+006
5	(R)RPCFSALEVDETYVPK(E)	637.96	1910.9322	1.02e+006
6	(K)YICENQDSISSK(L)	722.43	1443.6426	1.13e+006



Proteomics: Protein ID from yeast gel band tryptic digest.



Yeast proteome sample. The proteins from the yeast cell extract were separated by SDS-PAGE. Fifteen micrograms of protein was loaded in each lane of the gel. A single gel band was excised. Resulting peptides were isolated using the Agilent Protein In-Gel Tryptic Digestion Kit




Average number (triplicate) of identified peptides and proteins (Spectrum Mill) from the yeast gel band using the HPLC-Chip/MS(43 mm) versus conventional LC/MS with nanocolumns (43 mm and 150 mm).


- **4X more proteins were identified** with better sequence coverage for the tryptic digest of the yeast gel band with HPLC-Chip vs. comparable conventional nanocolumn.
- **Better chromatographic performance and MS sensitivity** than conventional nanocolumns.

Protein Biomarker Workflow: Protein Identification Approach

Identify proteins in each sample



Compare identified proteins across multiple samples



CHexaglycyl-depleted # spectra mean intensity	CHexaglycyl-enriched # spectra mean intensity	Protein MW (Da)	Database Accession #	%AA Coverage	Distinct Peptides (#)	Distinct Summed Score	Protein Name
196 2.81e+008	24 2.17e+007	5155753	7128	23	77	1128.83	1 epolipoprotein B-100 precursor
403 2.76e+008	38 1.81e+007	1871653	453783	42	65	1007.17	2 Complement C3 precursor [Contains: C3a anaphylatoxin]
785 2.67e+008	46 2.88e+007	163278	453722	46	49	787.50	3 Alpha-2-macroglobulin precursor (Alpha-2-M)
11 2.67e+008	19 8.17e+007	693671	453592	62	34	552.80	4 Serum albumin precursor
89 2.22e+008	19 3.37e+007	1927525	2671645	22	34	517.75	5 dJ34F7.4 (complement component 4A)
36 1.94e+008	124 5.81e+007	77050.4	453781	45	26	452.65	6 Serotransferrin precursor (Transferrin) (Siderophilin) (Beta-1-metal binding globulin)
31 2.07e+008	4 2.17e+007	248075.6	30722344	21	30	445.35	7 fibronectin 1 isoform 1 preproprotein; cold-insoluble globulin
13 2.24e+008	4 1.34e+007	1391263	4504075	32	29	400.24	8 Complement factor H precursor (H factor 1)
630 2.13e+008	154 6.33e+007	30778.0	453724	52	18	291.14	9 Apolipoprotein A-I precursor (Apo-AI)
19 1.81e+008	18 2.75e+007	53050.1	9845455	53	17	288.87	10 group-specific component (vitamin D binding protein); hDBP
87 2.07e+008	16 1.25e+007	1154723	1635009	23	16	278.92	11 ceruloplasmin
196 1.71e+008	64 6.75e+007	492957	133309	45	14	222.61	12 hemopexin
49 2.45e+008	4 4.36e+007	700373	4504035	31	13	213.39	13 Prothrombin precursor (Coagulation factor II)
137 1.32e+008	1 9.44e+006	90568.6	625234	31	15	203.04	14 plasmin (EC 3.4.21.7) precursor [validated]
6 6.06e+007	43 2.15e+007	45305.6	4826262	31	13	199.59	15 Heptoglobin-2 precursor
19 2.71e+008	4 7.57e+007	478835	4504892	32	14	195.18	16 kininogen, LMW precursor [validated]
11 2.03e+008	3 1.44e+007	453813	450213	34	13	191.31	17 epolipoprotein A-IV precursor
192 1.86e+008	183 1.83e+008	361522	12054072	50	12	183.27	18 immunoglobulin heavy chain constant region gamma 1
48 1.85e+008	2 9.35e+007	383125	4537202	44	10	178.20	19 apolipoprotein H precursor [validated]
14 1.81e+008	31 4.95e+007	685105	87019	26	12	160.89	20 Ig mu chain precursor; membrane-bound (clone 201)
54 2.13e+008	3 8.85e+007	158871	450772	65	8	147.98	21 Transferrin precursor (Prealbumin) (TBPA) (TTR) (ATTR)
10 2.82e+008	6 1.15e+007	51941.0	69990	22	9	145.41	22 alpha-1-B-glycoprotein
23 1.32e+008	2 3.47e+007	103373.0	4506040	16	9	139.73	23 inter-alpha-trypsin inhibitor family heavy chain-related protein
1 1.32e+008	3 3.47e+007	526027	4502261	29	10	137.46	24 Antithrombin-II precursor (ATIII) (PRO309)



Sample Preparation For Biofluids

Blood, Plasma, Cerebral Spinal Fluid, Urine

Sample
Preparation


A polyclonal antibody based system to rapidly deplete multiple high abundant proteins in serum/plasma/CSF.

- ➡ Launched in August 2003
- ➡ Individual Ab materials are mixed in selected percentages and packed into a column format
- ➡ Agilent continues to innovate and lead this market



Agilent Multiple Affinity Removal System: Where Are We? What is Next?

Sample
Preparation



Multiple Affinity Removal System 7

New depth with speed


Fully removes seven most abundant proteins in human plasma to unmask important otherwise undetectable proteins

- Highest reproducibility
- Lowest cost $\mu\text{L}/\text{sample}$
- Rapid methodology (30 minutes)
- Reusable for more than 200 experiment runs




Sample
Preparation

Orderable Now





2DGE- Before



2DGE- After

Multiple Affinity Removal System 14

Coming Soon!




Proteomic Sample



High Abundance Protein



Low Abundance Protein



Proteomic Sample with Low Abundance Proteins In Column Flow-through

Removes 92-94% of total protein

pI-based Fractionation: OFFGEL Principle

- after rehydration the IPG gel seals tightly against the well frame
- the diluted sample is distributed across all wells in the strip
- after fractionation, the liquid fractions are recovered with a pipette


Number of fractions: 12 or 24

Fraction volume: 150 μ L

Resolution: 0.1/0.6 pH

Loading capacity: 50 μ g – 5 mg per sample

Fractionation time: 8 - 24 h



OFFGEL Protein Fractionation

High *pI* resolution

- theoretical resolution 0.1 pH




E.coli lysate

OFFGEL Electrophoresis




2D-Gel
Electrophoresis

1D-Gel
Electrophoresis



pH 4-7, 24 fractions




pH 4-7, 12 fractions



OFFGEL Peptide Fractionation

E.Coli Tryptic Digest



=> 90% of peptides are found in 1 or 2 fractions!

**Number of OFFGEL fractions containing each individual peptide
(absolute numbers of peptides in parenthesis)**

Mass Profiling- Find The Differences Between Samples



Molecular Feature: a discrete molecular entity defined by the combination of

- retention time, mass and response in an LC/MS analysis
- retention time, mass spectrum and response in a GC/MS analysis

**LC/MS or
GC/MS
Analysis**

**Find
molecular
features**

**Compare
sample
sets**


Identification

Validation


- Find real differences in sample sets using statistical analysis
- Reproducible measurements minimize the number of samples!






Protein Biomarker Workflow: Protein Profiling Approach



MS Profiling And MS/MS Identification With The Agilent 6510 Q-TOF And HPLC-Chip



MassHunter Profiling Software




Mass Profiler




Importance of Informatics In Integrating Systems Biology

Genetics ➡ **Transcriptomics** ➡ **Proteomics** ➡ **Metabolomics**


Chrom. DNA/Genes




mRNA



Proteins



Metabolites




Informatics

Knowledge



GeneSpring Platform



Candidate Identification From Targeted MS/MS

Protein/Peptide Summary - Agilent Spectrum Mill - Microsoft Internet Explorer provided by Agilent Technologies, Inc.

Address: http://spectrummillid/millhtml/summaryframe.htm

Agilent Spectrum Mill - Protein/Peptide Summary

30min # spectra mean intensity	60min # spectra mean intensity	Archakov\SLX- fis # spectra mean intensity	Database Accession #	%AA Coverage	Distinct Peptides (#)	Summed MS/MS Search Score	Group #	Protein Name
4 1.11e+006	12 1.08e+007	28 1.24e+006	P20029	39	21	319.95	1.1	78 kDa glucose-regulated protein precursor (GRP 78) (Immunoglobulin heavy chain-binding protein) (BiP)
9 2.23e+006	14 1.02e+007	22 2.12e+006	P09103	38	18	261.60	2.1	Protein disulfide-isomerase precursor (EC 5.3.4.1) (PDI) (Prolyl 4-hydroxylase beta subunit) (Cellular thyroid hormone-binding protein)
9 3.26e+006	27 1.10e+007	33 1.72e+006	Q64458	39	14	215.05	3.1	Cytochrome P450 2C29 (EC 1.14.14.1) (CYP2C29) (P-450 MUT-2) (Aldehyde oxygenase)
6 1.86e+006	16 8.35e+006	12 2.50e+006	Q8VCT4	33	13	201.81	4.1	Carboxylesterase 3 precursor (EC 3.1.1.1) (Triacylglycerol hydrolase) (TGH)
2 3.76e+005	11 1.43e+007	19 2.40e+006	Q64459	31	13	196.62	5.1	Cytochrome P450 3A11 (EC 1.14.14.1) (CYP3A11) (P-450 3A11) (P-450UT)
10 3.66e+006	11 1.38e+007	20 1.76e+006	Q63880	34	13	195.98	6.1	Liver carboxylesterase 31 precursor (EC 3.1.1.1) (ES-Male) (Esterase-31)
3 6.26e+005	4 5.82e+006	14 1.23e+006	P37040	22	10	154.70	7.1	NADPH-cytochrome P450 reductase (EC 1.6.2.4) (CPR) (P450R)
1 7.15e+005	5 1.13e+007	12 8.57e+005	P27773	20	10	142.58	8.1	Protein disulfide-isomerase A3 precursor (EC 5.3.4.1) (Disulfide isomerase ER-60) (ERp60) (58 kDa microsomal protein) (p58)
5 2.96e+006	10 8.80e+006	9 1.81e+006	P24456	27	9	140.41	9.1	Cytochrome P450 2D10 (EC 1.14.14.1) (CYP2D10) (P450-16-alpha) (P450CB) (Testosterone 16-alpha hydroxylase)
7 4.17e+006	19 1.36e+007	21 1.73e+006	Q63886	25	9	136.88	10.1	UDP-glucuronosyltransferase 1-1 precursor, microsomal (EC 2.4.1.17) (UDPGT) (UGT1*1) (UGT1-01) (UGT1A1)
0 0.00e+000	4 7.76e+006	14 1.21e+006	Q90379	23	10			2.3) (Microsomal epoxide hydrolase) (Epoxide hydratase)
3 4.57e+006	5 1.11e+007	10 1.68e+006	Q62397	22	9			14.1) (CYP11B10) (Testosterone 16-alpha hydroxylase) (P450-16-alpha) (Clone PF3/46)
3 2.72e+005	6 1.05e+007	8 2.17e+006	P08113	13	9			mic reticulum protein 99) (94 kDa glucose-regulated protein) (GRP94) (ERp99) (Polymorphic)
5 7.29e+005	2 6.83e+006	3 1.06e+006	Q08601	11	8			protein large subunit precursor
9 8.90e+005	7 7.21e+006	14 7.13e+005	Q8JZQ0	16	7			ase 5 (EC 6.2.1.3) (Long-chain acyl-CoA synthetase 5) (LACS 5)
3 1.12e+006	2 6.06e+006	6 1.35e+006	Q05421	14	7			4.1) (CYP11E1) (P450-J) (P450-ALC)
0 0.00e+000	5 1.10e+007	4 1.20e+006	P08003	15	7			precursor (EC 5.3.4.1) (Protein ERp-72) (ERp72)

Spectrum Mill



Protein Biomarker Studies: Simple Model Mixture

Three mixtures prepared which contained:

- 100 fmol/ μ L BSA digest and
- either 25, 50 or 100 fmol/ μ L serotransferrin digest

Five replicate injections (0.2 μ L injection volume)


Rapid (9 min injection-to-injection) method used

Optimization of acquisition rate also investigated




Simple Model Study: Molecular Feature Extraction


Raw Data



Extracted Features




Simple Model Study: Finding The Differential Features in GeneSpring MS



- Filter data by relative frequency in replicates, fold change etc.
- Perform ANOVA to identify features that account for differences between sample groups
- PCA to visualize results shows clear differentiation between 3 sample groups
- Export list of m/z and retention time for targeted MS/MS

Simple Model Study: Protein Identification From Targeted MS/MS



Group (#)	Spectra (#)	Distinct Peptides (#)	Distinct Summed MS/MS Search Score	% AA Coverage	Mean Peptide Spectral Intensity	Database Accession #	Protein Name		
1	9	9	126.39	13	5.96e+005	Q29443	Serotransferrin precursor (Transferrin)		
#	Filename	z	Score	SPI (%)	Spectrum Intensity	Sequence	RT (min)	m/z Measured (Da)	MH ⁺ Matched (Da)
1	mix-C.1163.1169.2	2	17.75	93.6	7.19e+005	(K) CGLVPVLAENYK (S)	4.08	681.8701	1362.709
2	mix-C.1033.1047.2	2	18.53	96.2	7.83e+005	(K) GYLAVAVVK (K)	3.60	460.2938	919.561
3	mix-C.0795.0799.2	2	15.09	80.2	4.84e+005	(K) CLASIAK (K)	2.72	381.7190	762.418
4	mix-C.0887.0907.2	2	12.32	81.3	5.94e+005	(K) DSADGFLK (I)	3.06	426.7164	852.410
5	mix-C.0913.0931.2	2	14.80	85.6	4.29e+005	(K) ELPDPQESIQR (A)	3.16	656.3430	1311.654
6	mix-C.0917.0917.2	2	8.83	84.1	6.32e+004	(K) DQTVIQNTDGMNNEAWAK (N)	3.17	1009.4843	2017.921
7	mix-C.0956.0962.2	2	12.94	85.4	2.83e+005	(K) NYELLCGDNTR (K)	3.31	677.8183	1354.606
8	mix-C.0965.0973.2	2	9.89	68.0	1.63e+006	(K) ENFEVLCK (D)	3.35	519.7610	1038.493
9	mix-C.1021.1028.2	2	16.24	97.7	3.78e+005	(K) TSDANINWNLK (D)	3.55	695.3555	1389.676



Protein Biomarker Studies: Complex Model Mixture

Five mixtures prepared which contained:

- Tryptic digest of an *E. coli* lysate (complex background)
- Tryptic digests of bovine serotransferrin and BSA at specified levels

Each mixture was prepared 10 times

One injection of each sample (1 μ L injection volume)

Long (100 min injection-to-injection) method used




Complex Model Study: Spiked *E. coli* Lysate

Sample	<i>E. coli</i> lysate (ng total protein)	BSA (fmol)	Serotransferrin (fmol)
1	400	25	200
2	400	50	100
3	400	100	50
4	400	200	25
5	400	400	5



Complex Model Study: Spiked *E. coli* Lysate Total Ion Chromatogram




Digest of *E. coli* lysate creates a complex background of thousands of peptides


Profiling software must be able to find the bovine peptides that were spiked into the *E. coli* mixture

Complex Model Study: Molecular Feature Extraction


Raw Data



Extracted Features



Complex Model Study: Reproducibility With Technical Replicates



Color-coding by
molecular feature


Total of 762 features
shown

Average SD is
0.0519 min for RT
1.6 mDa for mass




Complex Model Study: Clustering of Features That Are Significantly Different (ANOVA)


Cluster by sample




Cluster by level



Complex Model Study: Finding The Differential Features in GeneSpring MS





Complex Model Study: K-Means Clustering of Differential Features And PMF Search



Complex Model Study: Mass Spectra From a Targeted Peptide


Apex mass spectrum for a targeted species

Fragment ion assignments for targeted peptide (from transferrin)



Post-translational Modifications (PTMs)

- Are chemical structures added to proteins **AFTER** they are translated
- **Control** or **regulate** the protein's structure or activity, such as ability to form a complex or interact with other proteins



Most Common Post-translational Modifications

Table A.1C.4 Mass Changes Due to Some Post-Translational Modifications of Peptides and Proteins^a

Modification ^b	Monoisotopic mass change	Average mass change
<i>Common modifications</i>		
Pyroglutamic acid formation from Gln	-17.0265	-17.0306
Disulfide bond (cystine) formation	-2.0157	-2.0159
C-terminal amide formation from Gly	-0.9840	-0.9847
Deamidation of Asn and Gln	-0.9840	-0.9847
Methylation	14.0157	14.0269
Hydroxylation	15.9949	15.9994
Oxidation of Met	15.9949	15.9994
Proteolysis of a single peptide bond	18.0106	18.0153
Formylation	27.9949	28.0104
Acetylation	42.0106	42.0373
Carboxylation of Asp and Glu	43.9898	44.0098
→ Phosphorylation	79.9663	79.9799
Sulfation	79.9568	80.0642
Cysteinylation	119.0041	119.1442
Glycosylation with pentoses (Ara, Rib, Xyl)	132.0423	132.1161
Glycosylation with deoxyhexoses (Fuc, Rha)	146.0579	146.1430
Glycosylation with hexosamines (GalN, GlcN)	161.0688	161.1577
→ Glycosylation with hexoses (Fru, Gal, Glc, Man)	162.0528	162.1424
Modification with lipoic acid (amide bond to lysine)	188.0330	188.3147
Glycosylation with <i>N</i> -acetylhexosamines (GalNAc, GlcNAc)	203.0794	203.1950
Farnesylation	204.1878	204.3556
Myristoylation	210.1984	210.3598



PTMs Are Dynamic In An Organism


The number of phospho groups on a phosphoprotein, or the number and type of glycans on a glycoprotein, change in response to changes in the environment

- Phosphate (sugar) concentration
- Enzyme concentration
- Cofactor concentration
- Nutrients present
- State (healthy or diseased, etc.)

In turn, the phospho- or glycoprotein interacts with other proteins (or enzymes, receptors...), triggering other actions



PTM Identification Workflow



Global Proteomic Profiling of Phosphopeptides Using HPLC-Chip Electron Transfer Dissociation Ion Trap Mass Spectrometer

- Identified 1435 unique phosphorylation sites from 500 proteins
- ETD identified 60% more phosphopeptides than CID, with an average of 40% more fragment ions
- Discovered 15 novel phosphorylation motifs



PNAS

Global proteomic profiling of phosphopeptides using electron transfer dissociation tandem mass spectrometry

Henrik Molina*[†], David M. Horn[‡], Ning Tang[‡], Suresh Mathivanan*[§], and Akhilesh Pandey*[†]

*McKusick-Nathans Institute for Genetic Medicine and Departments of Biological Chemistry, Pathology, and Oncology, The Johns Hopkins University School of Medicine, Baltimore, MD 21205; [†]Department of Biochemistry and Molecular Biology, University of Southern Denmark, Odense 5230, Denmark;

[‡]Agilent Technologies, Santa Clara, CA 95052; and [§]Institute of Bioinformatics, International Tech Park, Bangalore 560 066, India


Communicated by Paul Talalay, Johns Hopkins University School of Medicine, Baltimore, MD, December 25, 2006 (received for review November 1, 2006)

PNAS | February 13, 2007 | vol. 104 | no. 7 | 2199–2204




ETD Fragmentation

- Produces c and z fragment ions, while CID produces b and y fragment ions
- **Preserves post-translational modifications**
- ETD fragmentation is random and less sequence dependent



Instrument Schematic For ETD Ion Trap




ETD Advantages:

- Complementary fragmentation to CID by providing additional c and z ions
- Soft fragmentation process which maintains intact phospho-residues allows phosphorylation site determination
- Multiple MS(n) fragmentation methods with the combination of CID and ETD enrich information content
- Provides a new gas phase ion reactor for studying gas phase ion-ion chemistry reactions in real time




Experimental Design





Fractionation of Intact Proteins Using mRP Column

- Protein recovery > 98%
- Achieve high resolution
- Highly reproducible




Detailed Analysis of Phosphorylation Sites Identified Using ETD From the Lys-C digests

- 676 unique peptides were identified, derived from 287 proteins
- 606 (90%) were phosphorylated
- 344 singly phosphorylated, 208 doubly phosphorylated and 54 peptides contain three or more phosphorylation residues
- 84% serine phosphorylation, 14% threonine phosphorylation and 2.3% tyrosine phosphorylation




Phosphopeptides Identified by ETD and CID From LysC digest



ETD resulted ~40% more back-bone fragment ions transforming into a 23% better sequence coverage over CID


Charge State Distribution of the Identified Phosphopeptides



ETD identified more phosphopeptides with higher charge states, while CID identified more doubly charged precursors.


Phosphopeptides Identified by Three Proteolytic Digestion and CID/ETD

- ETD identified 60% more phosphopeptides than CID
- Trypsin and Lys-C identified similar number of phosphopeptides using ETD
- Glu-C identified fewer phosphopeptides than Trypsin and Lys-C




Correct Phosphorylation Location Determined by ETD

K/E t E S E A/E D/N/L/D/D/L E/K




K/E T E s E A/E D/N/L/D/D/L E/K




ETD Provided More Fragment Ions And Higher Sequence Coverage

E S E/D/K P E/I/E/D/V/G/s/D E E E E K K



E/S/E/D/K P E/I/E/D/V G/s/D E E E E K K

ETD



Discovery of New Phosphorylation Motifs

- 80% of identified phosphorylation sites were novel
- Identified 294 of 368 reported phosphorylation sites from 103 separated publications
- Of the 1141 novel phosphorylation sites, 85% matched known motifs
- 15 new motifs were identified by hierarchical clustering

	Novel motif	Occurrence in phosphopeptides
1	pS[E/D]X[E/D][E/D]	55
2	pSPXXXP	31
3	pSPXXXT	27
4	DXXXp[S/T]P	14
5	GGpS	13
6	p[S/T]PPP	12
7	QXp[S/T]P	12
8	PSp[S/T]P	11
9	PPXp[S/T]P	9
10	PPp[S/T]P	9
11	EXSXp[S/T]P	9
12	PXpSPX[R/K]	8
13	PpSXL	7
14	PLp[S/T]P	6
15	TpTP	5



Summary

- A complete solution for phosphoproteome analysis has been demonstrated
- 1435 unique phosphorylation sites from 500 proteins has been identified from human embryonic kidney cells
- ETD identified 60% more phosphopeptides than CID, with an average of 40% more fragment ions
- Discovered 15 novel phosphorylation motifs



Glycan Profiling

The majority of recombinant protein drugs used for treatment of disease are glycoproteins, parts of glycoproteins, or contain glycoproteins (examples: recombinant antibodies, vaccines)

The glycan portion of the product varies dramatically depending on the cell line used, fermentation conditions, and other production parameters, leading to batch-to-batch variations

The product quality and efficacy (e.g., circulation lifetime) varies dramatically depending on the identity and amount of the glycans present

One way of monitoring the quality of recombinant protein drugs is to profile the glycans, verifying that the product quality has not change appreciably



Some Product Names You May Recognize...


Product	Used to treat	Manufacturer
Activase® , TNKase® (t-PA, tissue plasminogen activator)	Heart attack (> 1 million US cases/ year), stroke (700,000 US cases/ year), pulmonary embolism (600,000 US cases/ year)	Genentech
Aranesp® (darbepoetin alfa)	Kidney failure, anemia	Amgen
Avastin® (monoclonal antibody)	Colorectal cancer	Genentech
Enbrel® (dimeric fusion protein)	Inflammation (e.g., rheumatoid arthritis)	Amgen
EPOGEN® (epoetin alfa)	Kidney failure, anemia	Amgen
Erbitux® (monoclonal antibody)	Colorectal cancer	ImClone Systems
Gardasil® (genetically engineered vaccine)	HPV infection leading to cervical cancer (10,000 US cases/ year)	Merck
Herceptin® (monoclonal antibody)	HER2-positive breast cancer	Genentech
Pediarix® (genetically engineered vaccine)	Diphtheria, pertussis, tetanus, hepatitis B, and polio in infants	GlaxoSmithKline
Raptiva® (monoclonal antibody)	Plaque psoriasis	Xoma, Genentech
Rituxan® (monoclonal antibody)	Non-Hodgkin's lymphoma	Genentech

All of these products

- are **glycoproteins** or contain glycoproteins produced by recombinant techniques
- have been approved by the FDA (commercially available) or are in the final stages of approval
- need to be checked for **stability** during their product development and phase 1 trials
- need to be checked for **consistency** during product development and phase 1 trials
- need ongoing QA/QC checks during production, especially of the glycan portion, to ensure
 - batch-to-batch uniformity of the product
 - consistent performance
 - they are **safe** for human consumption













Human Serum IgG N-Linked Glycan Alditols



Assuming a GlycoSuite database search segment score of 1, 2, or 4; species *Homo sapiens*, tissue *hemic system* or *immune system*; and attached protein is *IgG*, then these identifications are the only results that pass all criteria

Human Serum IgG N-Linked Glycan Distribution

Glycan (including isoforms)	Structure	<i>m/z</i> value of alditol $[M+2H]^{2+}$	peak height in averaged spectrum	relative percent
G0		733.4	46886	14.3
G1	 or 	814.4	103172	31.4
G0 + bisecting GlcNAc		835.2	30969	9.4
G2		895.4	35483	10.8
G1 + bisecting GlcNAc		916.0	25222	7.7
G2 + bisecting GlcNAc - fuc		924.0	7623	2.3
G1 + NeuAc		959.9	16101	4.9
G2 + bisecting GlcNAc		997.0	24573	7.5
G2 + NeuAc		1041.0	38886	11.8



Some inaccuracy in the relative amount is to be expected because of variations in response, charge state distribution, and adduct formation among the various analytes



Metabolite study: 30 μ M Buspirone Incubation

EIC m/z 360; 376

Narrow bore column



>100x
sensitivity


- Peaks 1 to 3
- Peaks 4

[M+H]⁺ = 376
[M+H]⁺ = 360

N,N-desethyl hydroxybuspirone
N,N-desethyl buspirone




Pharma: HPLC-Chip sensitivity, sulfadimethoxine in serum, 1 uL injection



Spiked plasma sample was centrifuged in a 3 kDa MW cutoff spin tubes. 1 uL is injected on to the enrichment column of the HPLC-Chip.

Enrichment column



Area RSD's of 2.6% (10 pg) to 10 % (10 fg) are obtained for 1 uL injection (without ISTD use).

Estimated LOD: 10 fg

- Automated sample enrichment at physiological concentration
- Small sample size: 1 μ L
- **125 X better sensitivity** that best reported QQQ data (1.25 pg/ μ L in milk)





Environmental: Residual herbicides in drinking water with automated sample pre-concentration.

- EU drinking water regulations require monitoring for neutral and acidic herbicides at the ppt level.
- Current procedures involving collecting from 1 to 5 liters of sample and then concentrating the herbicides using SPE prior to analysis.
- Detection using conventional 4.6 mm ID columns with LC/MS
 - Water sample spiked with herbicides at 1 pg/ μ L level.
 - Direct injection of 0.5 μ L of water with automated sample pre-concentration on HPLC-Chip (500 fg on-column).
 - Detection on 6340 Ion Trap MS




Environmental: Residual herbicides in drinking water with automated sample pre-concentration.

Agilent 6340 Ion Trap MS with HPLC-Chip



- Sensitivity enhancement of nanospray LC/MS allows detection at levels that meet EU drinking water regs (< 100 ppt) without SPE pre-concentration (reduced sample handling/preparation cost).
- HPLC-Chip enables routine, reliable use of nanospray LC/MS.



Michael Zumwalt and Hongfeng Yin
Agilent Technologies, Inc.


Diana Wilkins and Shannon McOmber
Center for Human Toxicology, Univ. Utah

ASMS 2007




MS Solution: Sensitivity and confirmation

QQQ - MRM (Multiple Reaction Monitoring)



HPLC-Chip/QQQ Technology – system configuration



Analysis of cTHC using HPLC-Chip/QQQ

LC conditions

❑ 1200 Series NanoLC

solvent A: 0.1% formic acid in water

solvent B: 0.1% formic acid in 95/10 acetonitrile/water

flow: 300 nL/min

Analysis of **standards**

isocratic: 30% B

Stop 5 min

❑ 1200 Series CapLC (2nd pump for sample loading on enrichment column)

1 µL inj. vol. (**standards**) 2 µL inj. vol. (**extracts**)

30:70 water/acetonitrile

flow: 3 µL/min

2 min load time

❑ HPLC-Chip/MS

Analytical column: 75 µm x 43 mm, 300 Å, 3.5 µm, C18

Analysis of cTHC using HPLC-Chip/QQQ – MS conditions

❑ Agilent G6410A QQQ Mass Spectrometer

Source (positive nanoelectrospray)

Capillary: 1800 V

Dry gas: 6 L/min

Gas temp.: 350 C



Analyzer

Monitored MRM transitions:

Cmpd	Transition	Dwell (msec)	Frag (V)	CE (V)
cTHC	345.1 > 299.2 (quant)	100	140	18
	345.1 > 327.2 (qual)			
d3-cTHC	348.1 > 302.2 (quant)	↓	↓	↓
	348.1 > 330.2 (qual)			




Increase in sensitivity over standard HPLC





30 x increase in signal


Injection of cTHC standard



Linearity and reproducibility of cTHC standard



Limit of quantitation in the low femtograms level




Replicate injections of
20 fg on-column

Area RSD = 12.5%

Average relative difference
between calculated and
expected concentrations
is only 7%

Results of standards – all transitions at lowest level

20 fg on-column



cTHC quant


cTHC qual

d3-cTHC quant



d3-cTHC qual

$\frac{1}{2}$ amount normally injected on-column by GC/MS/MS or GC/GC/MS


Results of standards – Compound Information



Linearity of cTHC in hair – >2 orders of magnitude



Limit of quantitation for cTHC in hair



400 fg on-column
spiked blank

Corresponds to 0.50 pg/mg
in hair with 100% extraction
recovery

Area RSD = 9%

Summary

- ❑ HPLC-Chip increases sensitivity by factor of 30 over standard LC flow
- ❑ Excellent sensitivity with standards, 20 fg on-column easily seen
- ❑ Excellent reproducibility at each level and $R^2 > 0.999$
- ❑ Good linearity of sample calibration curve, $R^2 > 0.998$, with excellent reproducibility at each level
- ❑ LOQ of 0.01 ng/mg in hair with RSD = 6%
- ❑ 400 fg matrix spike on-column (0.5 pg/mg) RSD of 9% shows sensitivity of LCMS using HPLC-Chip is 10x level achievable by GCMS without the need for derivatization. Note that this assumes 100% sample extract recovery. Results can be improved with better sample extraction.



Protein Quantitation with HPLC-Chip / MS and Stable Isotope Labeled Peptides




Christine Miller and Hongfeng Yin
Agilent Technologies, Inc.

ASMS 2007



Protein Quantitation Strategy



Information is kindly provided by Jon Gingrich of Sigma-Aldrich Co.

SIGMA-ALDRICH



Peptide Selection (myoglobin)


ALELFR

148-153

AQUA peptide

HPLC-Chip/MS

MRM Transitions for Peptide ALELFR



Quantitation of Myoglobin Digest Spiked in Digested Human Serum

calibration level	Myoglobin		Internal standard	Serum	water
	final concentration	added amount			
1	0.4fmol/μL	2μL*20fmol/μL	10μL	2μL	86μL
2	1fmol/μL	5μL*20fmol/μL	10μL	2μL	83μL
3	2fmol/μL	10μL*20fmol/μL	10μL	2μL	78μL
4	5fmol/μL	25μL*20fmol/μL	10μL	2μL	63μL
5	10fmol/μL	50μL*20fmol/μL	10μL	2μL	38μL
6	20fmol/μL	13.5μL stock	10μL	2μL	74.5μL
7	50fmol/μL	33.5μL stock	10μL	2μL	54.5μL
8	100fmol/μL	67μL stock	10μL	2μL	21μL

Myoglobin Stock Concentration is 150fmol/μL

Internal Standard Stock Concentration is 10fmol/μL LALELFR* and 100fmol/μL LFTGHPETLEK*


Final Internal Standard Concentration is 1fmol/μL LALELFR* and 10fmol/μL LFTGHPETLEK*

Serum Concentration is 12mg/mL. Final concentration is 240ng/μL

Aqua peptides were kindly provided by Jeffrey Porter and Jon Gingrich of Sigma-Aldrich Co.



Detection Sensitivity of the Native Myoglobin Peptide with HPLC-Chip/MS




HPLC Chip: 43mm Zorbax SB C18
 nanoPump: 400nL/min, A. water+0.1%FA, B. 95%ACN 0.1%FA
 0 min 5%B, 10min 30%B, 12min 80%B
 CapPump: 3μL/min, 2%ACN water
 Sample: 0.4fmol/μL * 0.2μL injection, sample is in 240ng/μL
 digested serum.

Pept i de	pr ecur sor	f r a g m e n t	quant / qual
LFTG-PETLEK	636. 3	716. 4	quant
LFTG-PETLEK	636. 3	1011. 6	qual
LFTG-PETLEK*	640. 3	724. 4	quant
LFTG-PETLEK*	640. 3	1019. 9	qual
ALELFR	374. 7	564. 3	quant
ALELFR	374. 7	435. 3	qual
ALELFR*	379. 7	574. 3	quant
ALELFR*	379. 7	445. 3	qual




Calibration Curve for ALELFR Using ALELFR* as Internal Standard



Triplicate injections at each level

Detection Sensitivity with HPLC-Chip/QQQ




HPLC Chip: 43mm Zorbax SB C18
 nanoPump: 400nL/min, A. water+0.1%FA, B. 95%ACN 0.1%FA
 0 min 5%B, 10min 30%B, 12min 80%B
 CapPump: 3μL/min, 2%ACN water
 Sample: 0.04fmol/μL * 1μL injection, sample is in 240ng/μL
 digested serum

Pept i de	pr ecur sor	f r a g m e n t	quant / qual
LFTG-PETLEK	636. 3	716. 4	quant
LFTG-PETLEK	636. 3	1011. 6	qual
LFTG-PETLEK*	640. 3	724. 4	quant
LFTG-PETLEK*	640. 3	1019. 9	qual
ALEFLR	374. 7	564. 3	quant
ALEFLR	374. 7	435. 3	qual
ALEFLR*	379. 7	574. 3	quant
ALEFLR*	379. 7	445. 3	qual



Quantitation with Internal Standard







Peptide	precursor	fragment	quant / qual
LFTG-PETLEK	636.3	716.4	quant
LFTG-PETLEK	636.3	1011.6	qual
LFTG-PETLEK*	640.3	724.4	quant
LFTG-PETLEK*	640.3	1019.9	qual
ALEFLR	374.7	564.3	quant
ALEFLR	374.7	435.3	qual
ALEFLR*	379.7	574.3	quant
ALEFLR*	379.7	445.3	qual

The ALEFLR peak from 80amol digested Myoglobin is calculated to be 63 amol using the 40amol ALEFLR* Aqua peptide. One reason for it could be that the digestion efficiency is 80%.

Serum Protein Quantitation using Aqua Peptides

Aqua Peptides from PRS proteomics

Peptide	sequence	Aqua peptide	fragment	native peptide	native peptide fragment
P1	ESYSGVTL*DPR	615.8++	272 (13eV)	612.3 ++	601.3 (y5 ion singly charged) and 272 (y2)
P2	DIPTNSPELEETLTHTITK*	716.5+++	640.3 (9eV)	713.7+++	637.6 (y17 triply charged)
P4	TNLESILSYPK*	636.9++	815 (17eV)	632.9 ++	807.5 (y7 singly charged)
P6	SLDFTELDVAEEK*	723.4++	882.1 (18eV)	719.4 ++	874.5 (y ion singly charged)



Samples provided by Richard Jones of Proteomic Research Services, Inc.



Conclusions:

- ❑ High sensitivity peptide analysis using HPLC-Chip/QTOF and HPLC-Chip/QQQ MS
- ❑ High efficiency chromatography separation with HPLC-Chip
- ❑ Good retention time and MS detection reproducibility
- ❑ Excellent quantitation when using stable isotope labeled peptides as internal standard
- ❑ Sample loading capacity is at least 300ng total protein digests on a standard short analytical column/40nL enrichment HPLC-Chip



HPLC-Chip portfolio

- **MS Calibration & Diagnostic chip:** Easy infusion of MS tuning mix, Calibration of the Chip Cube valve
- **Protein ID Chip#1:** 43 mm column, peptide separation. Low/medium complexity tryptic digest mixtures. Dilute samples 100-400 ng on-column
- **Protein ID Chip # 2:** 150 mm column, peptide separation, medium/high complexity tryptic digest mixtures. Concentrated samples 500 ng to 1 µg on column.
- **Graphitized Carbon Chip:** Glycan (oligosaccharides) separations, 43 mm column. Highly polar compounds normally not retained on C-18 material. Structurally related compounds such as geometric isomers and diastereoisomers
- **Small molecule Chip:** Metabolite ID, DMPK, dilute samples. Small molecules retained and well characterized by SB-C18. 43 mm column
- **Infusion Chip:** direct infusion or automated flow injection of the sample directly in to the mass spectrometer at nanoflow rates for the collection of MS and MS/MS data.
- **SPQ Chips:** high capacity enrichment column, or intact protein separation chip using C-8 material
- **Custom Chip:** Custom packing, enrichment and/or analytical column.



HPLC-Chip Interface: Part of the Agilent LC/MS platform.



6300 Series Ion Trap MS



6500 Series QTOF MS



6100 MSD: Fall 2007



6200 Series TOF



1200 Series Nano-LC



6400 Series QQQ: June 2007



HPLC-Chip/MS: Conclusions



HPLC-Chip/MS Features:

- Increased productivity: easy to use, fast system setup and optimization, integrated workflows.
- Increased uptime: robust, reliable, plug and play operation designed for reliability and supportability
- Enhanced performance: improved MS/MS data quality, high sensitivity and high resolution chromatographic separations.
- Flexibility: optimized application and system performance simply by changing HPLC-Chip.

Customer feedback:

“The Chip changed my life”

“I have been able to triple the throughput in my lab!”

“Once you use the chip, you don’t want to go back to needles!”

“My technician can now run nano-LC/MS!”

“So reliable, I now run nanoLC sequences over the weekend!”

.....



Thank you very much for your attention!

