

BVT Technologies, a.s.

Specialized in:

- design and manufacturing of miniaturised TFT electrochemical sensor and biosensor substrates
- further products and accessories integrating the sensor response to user-communicable results, data acquisition systems, miniaturised pumps, microflow system, special connectors, adapters etc.



Characteristic of the company

- Established in 2000
- SME, up to 10 employees + external workers
- Equipped to produce screen printed electrochemical sensors and biosensors and accessories
- Staff with more than 20 years expertise in R&D of electrochemical sensors and biosensors
- Close co-operation with R&D centres (universities and research institutions on international basis)

Principle products



- **Sensor substrates:** amperometric
conductometric
- **Microflow system:** to remove conventional problems rising within the standard electrochemical measurement setup
- **Biosensor analyzer:** to record directly the sensor measurements and download the information for analysis on PC
- **Peristaltic pump:** micro-flow pump with changable flow-rate parameters controlled via PC
- **Manual screen printer:** for in-house small scale production of a wide variety of sensors.

Historie

- Výzkumný ústav zdravotnické techniky
- Laboratorní přístroje Praha (Dr. Vozka, Dr. Litomisky)
- Masarykova universita v Brně (UJEP) (Prof. Macholán, Doc. Skládal)
- Excellent Lanškroun (Ing. Burša)
- Endokrinologický ústav v Lubochni (Dr.A. Pírek, Dr. A.Kreze)
- MyLab Prešov (Ing.Chudik,..)
- VUEK –(Ing. Kurtev)
- Mesit – (Dr. Šuranský, Ing. Řezniček)


Příklady







PACI, 6.6.2007, Prague

Princip



Substrates for electrochemical biosensors



PACI, 6.6.2007, Prague

Amperometric sensor

Type: AC1.W*.R*(*)

Usage in measurement of:


- Basic electrochemical and bio-electrochemical techniques
- H₂O₂ concentration
- Glucose
- Toxicity caused by pesticides
- Enzyme activity
- Ferro – Ferricyanide couple

Weight: 0.4 gms *L-W-T:* 25,4x7.26x0.63mm

Working electrode material: Au, Pt, Ag, graphite, alloy of Au and Pt

Reference electrode material: Ag, AgCl, Ag covered with AgCl

AC2.W2.RS




PACI, 6.6.2007, Prague

AC3, AC4




PACI, 6.6.2007, Prague

AC5,AC8




PACI, 6.6.2007, Prague

AC9,AC10





AC6,AC12,AC11



PACI, 6.6.2007, Prague

CC1,CC2



PACI, 6.6.2007, Prague

Accessories

- Pumps
- Evaluating units
- Connectors
- Flow adapters
- Screen printer
- MFS

Pump



PACI, 6.6.2007, Prague

Bioanalyzer

BEEP, OFBio




PACI, 6.6.2007, Prague


Connectors



Flow adapters



Screen printer



PACI, 6.6.2007, Prague

Glass vessels




PACI, 6.6.2007, Prague

MFS – Microfluidic system




Software

The analysis of reaction kinetics




Software

The automatic analysis of electrochemical sensors response



Software

The software developed on demand of customer



PACI, 6.6.2007, Prague

Ukazka speciálního software

..\..\Programy\Syncdet\7428-SyncDet.exe

..\..\Programy\Modely\rozhrani.exe

..\..\Programy\7429-2BIOAN\Bioanalyzer.exe

Aplikace a odběry

Datum aplikace	Aplikované pesticidy
19.4.2004	AFALON 45 SC
27.5.2004	METHANION 48 EM
16.6.2004	METHANION 48 EM
28.6.2004	DITHANE M 45
16.7.2004	METHANION + DITHANE
6.8.2004	METHANION + DITHANE




Odběry vzorků

- 1 den před, v den postřiku, 1, 3 a 5 dnů po
- půda - sondovací tyč do hloubky 10 cm, 10 x,
- listy – náhodně 10 g z parcely
- hlízy – 2-3 ks z parcely

Universal Rotating Disc Electrode


Jan Krejčí⁽¹⁾, Petr Skládal⁽²⁾, Dagmar Krejcová⁽³⁾, J. Kubka⁽⁴⁾,
Z. Grosmanová⁽³⁾

- (1) Krejčí Engineering, Tišnov, Czech Republic
- (2) MU Brno, Czech Republic
- (3) BVT Technologies a.s., Hudcova 72c, Brno
- (4) EMO, Hudcova 72c, Brno




PACI, 6.6.2007, Prague



Arrays of microelectrodes



Direct voltametry on electrodes array





- The principle




Test of electrode arrays

- Very poor reproducibility. The signal changed more than 30 % in same condition only by new inserting of sensor to the testing chamber. The sensor is extremely sensitive to hydrodynamic conditions
- High cross talk between electrodes

Radial flow to electrode arrays



The principle of Rotating disc electrode



(a) Vector representation of fluid velocities near disk. (b) Schematic resultant streamlines (or flows).


Rotating disc electrode

Advantages


- The hydrodynamics is solved
- The response is known (Levich equation)
- The methods which enables the monitoring of chemical reaction kinetics are described

Technical solution


1. The solution is inserting through the axis of rotating element
2. The element rotates above array of electrodes



The flow characteristic




Miniaturization to compact device




Results

zones of 0.25 mM ferricyanide, 1 ml/min, -100 mV/int. ref. better mass transfer with stirring -
15-times enhanced response




Effect of velocity of stirring element



The test in FIA mode


peaks of 1 mM ferricyanide, 1 ml/min, -100 mV/int.
ref.

excellent reproducibility




Cross-Talk between electrodes

The result of measurement with immobilised glucose oxidase. The enzyme was immobilised on electrode no. 2,4,6,8. The electrodes 1,3,5,7 was without enzyme. The electrode was immersed in stirred vessel.



Cross-talk between electrodes

The result of measurement with immobilised glucose oxidase. The enzyme was immobilised on electrode no. 2,4,6,8. The electrodes 1,3,5,7 was without enzyme. The measurement was done with URDE cell.




Acknowledgement

This work was partially supported by

- EU project GRD1-2001-4183 Microprotein
- Project Antope FD-K2-53, Czech Ministry of Industry

The Origins of TFT




PACI, 6.6.2007, Prague

3D Thick Film Technology


Dr. Jan Krejčí

Principles of the TFT Electronic Circuit




PACI, 6.6.2007, Prague

The Chemical Circuit





Production principles of the 3D TFT Chemical Chip





PACI, 6.6.2007, Prague

Channel preparation of the 3D - TFT



Alumina substrate



Channel preparation of the 3D - TFT



Printing of the first layer



PACI, 6.6.2007, Prague

Channel preparation of the 3D - TFT




Channel height is created by repeated printing
PACI, 6.6.2007, Prague

Channel preparation of the 3D - TFT



The supporting membrane is inserted
PACI, 6.6.2007, Prague



Channel preparation of the 3D - TFT



The structure of the supporting membrane ensures that the paste material penetrates to the membrane



PACI, 6.6.2007, Prague

Channel preparation of the 3D - TFT




The last print finalises the channel

Channel preparation of the 3D - TFT



The paste penetrates to the membrane and the properties of the channel walls are determined by the printed paste material


Amperometric sensor with an exact reference electrode





An example of 3D-TFT use showing an amperometric sensor with an exact reference electrode.

PACI, 6.6.2007, Prague

Amperometric sensor with an exact reference electrode




Amperometric sensor with an exact reference electrode



Final product. The electrochemical sensor is 25 mm long and 7.3 mm wide. The diameter of the working electrode is 1 mm.


Ion-selective Electrode



The sensor consists of two potentiometric electrodes. The first one is a reference electrode and the second one is ion-selective.

PACI, 6.6.2007, Prague


Ion-selective Electrode



Ion-selective Electrode




Oxygen Electrode




The internal space is filled by an internal electrolyte. Small windows within the membrane enable oxygen diffusion to the working electrode.

Micro-filter



PACI, 6.6.2007, Prague


Valve



The supporting membrane can be made from metal. The membrane must be selectively cut by laser.


PACI, 6.6.2007, Prague

Valve




PACI, 6.6.2007, Prague

Capillary Electrophoresis



Capillary Electrophoresis



Final product

PACI, 6.6.2007, Prague





Fig. 5. Separation of Na, K and Li (as chlorides) in 20 mM MES-20 mmol His, pH 5.5; electrokinetic injection: 200 V/5 s; separation voltage: 600 V. Cathode on the right-hand-side.


Deyl Z., Rohlicek V.: Versatile tool for the manipulation of electrophoresis chips, *Journal of Chromatography B*, 770 (2002) 19-23

Capillary Electrophoresis




PACI, 6.6.2007, Prague

Accelerometer




PACI, 6.6.2007, Prague

Real structure of the channel



Channel Imprecision



Screen structure is only partially transferred to the printed object

PACI, 6.6.2007, Prague

Types of 3D Structures

Optimum structures for 3D –TFT

- Characteristic width
500 – 100
microns
- Characteristic height
10 – 500
microns
- Structure function
does not depend on
wall surface quality


Optimum structures for other technologies

- Characteristic width
less than 20
microns
- Characteristic height
less than 10
microns
- Structure function
depends on wall
surface quality

The Future

- New TFT Technologies (FODEL, Fine line printing, Photolithography of printed layers)
- Connection with LTCC
- Lab on a Chip


An example of connecting a capillary with LTCC



A 180 micron stainless steel tube is co-fired to LTCC.

PACI, 6.6.2007, Prague


3D-TFT Microfluidic module




Evaluating electronics, amplifiers,...

Sensor made on silicon

Microfluidics made by nanotechnologies
in silicon






Acknowledgements

This work was partially supported by:

- EU project GRD1-2001-4183 Microprotein
- Program Technos TC 5 – 119, Czech Ministry of Industry
- Project Antope FD-K2-53, Czech Ministry of Industry



Electrochemical
system using oxygen
evolving PSII
particles from the
cyanobacterium
*Synechococcus
elongatus*




Nanostructured Electrodes


- Krejčí J.#, Maly J., Jakubka L., Pilloton R., Sameh K., Steffan P., Sugiura M.

Princip






PACI, 6.6.2007, Prague




PACI, 6.6.2007, Prague

Průběh proudu senzorem v závislosti na koncentraci



PACI, 6.6.2007, Prague

Array of nanoelectrodes




PACI, 6.6.2007, Prague

Electrochemical system using oxygen evolving PSII particles from the cyanobacterium *Synechococcus elongatus*

Principle


Electron transport from PSII with functional OEC complex to the platinum working electrode mediated using the artificial electron acceptor duroquinone (DQ). Reduced duroquinone is reoxidised at +620mV (vs. Ag/AgCl reference electrode). Oxidation current is proportional to the electron transport activity of PSII. Electric current generated by PSII is lower in the presence of herbicide. Screen-printed Pt:Ag/AgCl electrodes in three electrode arrangements are used.

PACI, 6.6.2007, Prague



Synthesis of self assembled monolayer of Ni-NTA chelator on gold electrode

(i) selfassembled monolayer (SAM) of cysteamin is deposited on gold electrode, modified (ii) with glutaraldehyde and finally with nitrilotriacetic acid (NTA) (iii) (Maly J. et al. 2002).




The mass transport is influenced by immobilisation strategy of enzyme

Panel A:

The inner structure of BSA-GA gel matrix with immobilized photosystem II.


Panel B:

The structure of the oriented monolayer of photosystem II.



Assembling of PSII on gold electrode monitored on-line during process


Electric current response of the electrode monitored on-line during the loading of electrode with His-tagged PSII



Influence of spacer BSA biomolecule in monolayer of PSII on biosensor activity


Panel A:

The structure of oriented monolayer of PSII mixed with the BSA spacer




Panel B:

The structure of electrode with the nano-structured active surface.




Influence of spacer BSA biomolecule in monolayer of PSII on biosensor activity

Signal obtained during the time-course of measurement from electrode with the monolayer of His-tagged PSII mixed with BSA (**Panel A**) and with the monolayer of pure His-tagged PSII (**Panel B**).




Nanostructured architecture of electrodes for improvement of mediator diffusion

The proposed parameters of the electrode with nano-structured active surface.




Comparison of time dependent current densities obtained by simulation for spherical and planar electrode


- 1: R = 50 nm
2: R = 100 nm
3: R = 500 nm
4: R = 1 μm
5: R = 5 μm
6: planar electrode
R: radius of the nanoelectrode



Comparison of the current response of planar (2) and nanostructured (1) electrode

The surface of macroelectrode is 1 mm^2 . It consists of electrodes with diameter $1 \mu\text{m}$ and distance $2 \mu\text{m}$. The time window optimal for increasing the measured output signal is marked.





E. Huang, M. Satjapipat, S. B. Han, and F. M. Zhou, *Langmuir* 17:1215 (2001).

Acknowledgments

This work was supported by the EU project GROWTH GRD1 - 2001 - 41831 "Microprotein",

by the project 522/03/0659 of the Grant Agency of the Czech Republic and

by COSMIC Project (ENEA Target Project on Biosensors and Bioelectronics).

A few of our clientele

- **Biotechnology Division, ENEA C.R. Casaccia, Rome, Italy**
- **Cranfield Biotechnology Centre, Cranfield University, UK**
- **Chemical Engineering Dept., University Rovira I Virgili, Spain**
- **Department of Technological Chemistry, Universidade Federal do Maranhao, Brazil**
- **Department of Biochemistry, Masaryk University, Brno, Czech Republic**
- **Department of Analytical Chemistry, The Academy of Sciences of the Czech Republic (ASCR), Prague, Czech Republic**
- **Department of Analytical Chemistry, University of Pardubice, Czech Republic**
- **Dipartimento di Chimica, Bari, Italy**
- **Dipartimento di Sanita Publica, Epidemiologia e Chimica Analitica, Universita degli studi Firenze, Italy**
- **Department of Analytical Chemistry, University of Alcala, Spain**
- **Departamento de Quimica Fisica, Universidad de Salamanca, Spain**
- **Department de Quimica Ambiental, CSIC, Barcelona, Spain**
- **Department of Clinical Biochemistry, University of Newcastle, UK**
- **Institut fur Biochemie & Molekulare Physiologie, Universitat Potsdam, Germany**
- **Istituto di Chimica Nucleare, CNR, Rome, Italy**
- **Istituto Biochimica ed Ecofisiologia Vegetali, CNR, Rome, Italy**
- **Military Institute of Technology, Rybkova, Brno, Czech Republic**
- **Microbiology Institute, The Academy of Sciences of the Czech Republic (ASCR), Třeboň, Czech Republic**

WELCOME

BVT Technologies, a.s.

Hudcova 78

612 00 Brno

Czech Republic

www.bvt.cz

Tel/Fax No.: + 420 541 513 665

e-mail : info@bvt.cz



Contact Persons:

Oldriska Mojzisova

Dr. Jan Krejčí, Company Manager

Ing. Zuzana Grosmanová

THANK YOU FOR YOUR ATTENTION !